

Lab. ID #
118563
To: Town of Sherborn
 19 Washington St.
 Sherborn, MA 01770

Received Date/Time:	8/4/25, 1:00 PM
Analyzed Date/Time:	8/4/25, 1:45 PM
Reported Date:	8/5/25
Collected By:	G & L Labs.
Collection Date:	8/4/25

ANALYTICAL RESULTS

Sample#	Sampling Location	Sampling Time	E. coli (CFU/100mL)
1	Farm Pond	10:30 AM	20
Maximum Contaminant Level			235
Method of Reference			EPA 1603.1

Laboratory QC Results

QC	Unit	E. coli
Duplicate (D1)	CFU/100mL	20
Duplicate (D2)	CFU/100mL	12
Log Dif. of Dup		0.2218
MS Recovery	%	93%
MS Rec. Criteria	%	17%-117%
Dilution Blk.(Beginning)	CFU/100mL	Negative
Dilution Blk.(Ending)	CFU/100mL	Negative
Positive Control	CFU/100mL	Positive

Approved By:



Lab. Director

8/6/25

Date

Lab. ID# 118563

Revised 2018

Town/City: Sherborn, MA

Town/City:

Sherborn
08/04/85

Sherborn
08/04/05
P.S. G+L LBB

Time Delivered to Lab: 1:00 PM
Delivered by: P.D.
Relinquished to: SM

Relinquished to: SM

Cloud Cover: ☐ Clear ☒ Partly Cloudy ☐ Cloudy/Overcast

Comments:

A copy of this form must be included with the samples when they are submitted to the lab.